

## OptiPro Multi-Enzyme

Concentrated multi-enzymatic manual detergent concentrate for manual and ultrasonic cleaning

ARTG 235775



## Features

- A concentrated multi-enzymatic detergent with outstanding cleaning performance for manual and ultrasonic cleaning of surgical instruments and manual cleaning of endoscopes. Provides improved removal of hard-to-clean soils, reducing soaking time and brushing.
- Optipro is a concentrated multi-enzymatic detergent with outstanding cleaning performance for manual and ultrasonic cleaning of surgical instruments and manual cleaning of endoscopes.

### Outstanding Performance


- Advanced, multi-enzymatic technology reduces the soaking and scrubbing needed to remove surgical soils, improving process efficiencies and user experience.
- Versatile and multi-purpose – can be used as a pre-soak, manual cleaner, ultrasonic detergent, and flexible endoscope detergent in all water and soil conditions.
- Low-foaming, crystal clear use solution allows visibility of instrumentation in the sink.
- Neutral pH, non-corrosive and free-rinsing formulation is compatible with surgical instruments and flexible endoscopes.
- Effectively removes hard-to-see soil residues that can remain after cleaning.

### Advanced Technology

- Utilizes the latest enzyme technology specific for surgical soils.
- Optimized surfactants for removal of lipids.
- Suitable for use in a wide range of water temperatures.

### Sustainable Solution

- 2X concentrated liquid provides more uses per litre.
- Phosphate-free to help reduce environmental impact.
- Contains biodegradable surfactants.

Product Code	Product Description	Pack Size	Quote
71006911	OptiPro Multi-Enzyme Manual Concentrate Deterget, 3.78 L	4 Pk	

## Safety Information

---

Do not mix with any other products. For further information, safety instructions and health warnings, see product label and Safety Data Sheet.

ALWAYS READ THE LABEL. FOLLOW DIRECTIONS OF USE.

## Directions For Use

---

### Manual Cleaning of Flexible Endoscopes:

- The cleaning solution should be made fresh for every scope.
- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4 mL/L of warm water.
- Proceed with the flexible endoscope cleaning process per the device manufacturer's written IFU.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

### Ultrasonic Cleaning:

- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4mL/L of warm water.
- Use solution temperature should not exceed 60°C.
- Complete a degas cycle prior to cleaning.
- Run appropriate ultrasonic cleaning cycle per the device manufacturer's written IFU.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

### Automated Flushing of Flexible Endoscopes:

- The cleaning solution should be made fresh for every scope.
- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4 mL/L of warm water.
- Proceed with the automated flushing of flexible endoscope per the device manufacturer's written IFU.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

### Pre-Soak and Manual Cleaning of Medical Instruments:

- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4 mL/L of warm water.
- Use solution temperature should not exceed 43°C.
- Pre-soak instruments as needed until soil is dissolved and removed from the instruments. Soak for a minimum of 1-5 minutes.
- Additional soak time or light mechanical cleaning with an instrument brush may be necessary depending on instrument complexity or age of soil. Refer to IFU for brushing and cleaning.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

## Relevant Product

---

- |   |                                    |
|---|------------------------------------|
| • <a href="#">Soluscope Sprint</a>            | • <a href="#">Bandelin Sonorex</a> |
| • <a href="#">Soluscope Sprint PT</a>         | • <a href="#">Bandelin Sonomic</a> |
| • <a href="#">EndoSelect Cleaning Brushes</a> | • <a href="#">Bandelin Trison</a>  |

- [Belimed UW565 Ultrasonic](#)
- [Bandelin Sonoboard](#)

## Applications

---

- Infection Control
- Healthcare Facilities
- Central Sterile Departments
- Endoscopy
- Residential Aged Care

## Resources

---

- safety data sheet (sds)
- Video
- Brochure